

Inappropriate Off-label Use of a Qualitative, Point-of-care hCG Device

Robert D. Nerenz, PhD*
Ann M. Gronowski, PhD†
David G. Grenache, PhD‡

*Dartmouth-Hitchcock Medical Center, Department of Pathology and Laboratory Medicine, Lebanon, New Hampshire

†Washington University School of Medicine, Department of Pathology and Immunology, St. Louis, Missouri

‡University of Utah School of Medicine, Department of Pathology, Salt Lake City, Utah

Section Editor: Mark I. Langdorf, MD, MHPE

Submission history: Submitted September 28, 2016; Accepted October 11, 2016

Electronically published December 6, 2016

Full text available through open access at http://escholarship.org/uc/uciem_westjem

DOI: 10.5811/westjem.2016.10.32675

[West J Emerg Med. 2017;18(2)324-325.]

Dear Editor,

We read with interest the recent study by Gottlieb et al¹ describing the reduction in turnaround time achieved by substituting whole blood for urine on a qualitative point-of-care (POC) hCG device. The device used in this study is FDA-approved and CLIA-waived only when the manufacturer's instructions are followed: three drops of urine or serum are applied to the device and results are recorded within three minutes (urine) or five minutes (serum) after application of sample.² However, the practice described by the authors differs considerably from the manufacturer's instructions, as whole blood was used rather than urine or serum and results were interpreted after 10 minutes. Modification of an approved device constitutes off-label use, is considered a laboratory-developed test and requires extensive validation to establish the modified device's performance characteristics before it is used in a clinical setting.

We commend the authors for noting that qualitative POC hCG devices are not FDA-approved for use with whole blood and we acknowledge their concurrent testing of urine on the same POC hCG device as a reference method. However, in addition to a method comparison study, CMS requires that laboratory-developed tests undergo an evaluation of precision, analytical sensitivity, analytical specificity, reportable range, reference interval and any other pertinent performance characteristics prior to being released for clinical use.³ Although a method comparison was performed, many additional device performance characteristics have not been defined. Furthermore, validation study results are limited to the specific clinical setting in which the study was performed and are not transferable to another institution, meaning that each institution that intends to offer a laboratory-developed test for clinical use must perform its own validation study. Use of an uncharacterized device to make clinical decisions puts patients at risk for adverse outcomes, particularly if inappropriate treatment is administered to a

pregnant patient, an ectopic pregnancy goes undiagnosed due to a false negative result, or if necessary surgical intervention is delayed because of a false positive result. Use of modified devices without the required validation studies also jeopardizes the hospital laboratory's accreditation and may result in forced discontinuation of laboratory testing, which negatively impacts patient care throughout the hospital.

We support the authors' assertion that an FDA-approved device capable of rapid hCG detection in a whole blood specimen at the point of care would be valuable in healthcare delivery settings. We would like to point out that two FDA-approved test platforms are already available for exactly that: the Abbott i-STAT β hCG cartridge and the NowDiagnostics ADEXUSDx hCG test. In addition to receiving FDA approval, the performance characteristics of both of these devices have been independently evaluated in academic medical centers.^{4,5} We strongly recommend that the authors engage with laboratory professionals at their institution to discuss available testing options and select appropriate test methods that meet the clinical need without jeopardizing patient care.

Address for Correspondence: Robert D. Nerenz, PhD, Dartmouth-Hitchcock Medical Center, Department of Pathology and Laboratory Medicine, 1 Medical Center Drive, Lebanon, NH 03756. Email: Robert.D.Nerenz@hitchcock.org.

Conflicts of Interest: By the WestJEM article submission agreement, all authors are required to disclose all affiliations, funding sources and financial or management relationships that could be perceived as potential sources of bias. Ann M. Gronowski disclosed consultation from the Church and Dwight Co., Inc., and research support from Abbott Diagnostics & Abbott Point of Care Scientific & Medical Advisor Board: Theranos. David G. Grenache disclosed research support from paid speaker: Abbott Point-of-Care.

Copyright: © 2017 Nerenz et al. This is an open access article distributed in accordance with the terms of the Creative Commons Attribution (CC BY 4.0) License. See: <http://creativecommons.org/licenses/by/4.0/>

REFERENCES

1. Gottlieb M, Wnek K, Moskoff J, et al. Comparison of result times between urine and whole blood point-of-care pregnancy testing. *West J Emerg Med.* 2016;17:449-53.
2. Beckman Coulter Icon 25 hCG package insert, 2014.
3. U.S. Government Publishing Office. Available at: www.gpo.gov. 42 CFR 493.1253(b)(2).
4. Sowder AM, Yarbrough ML, Nerenz RD, et al. Analytical performance evaluation of the i-STAT total β -human chorionic gonadotropin immunoassay. *Clin Chim Acta.* 2015;15:165-70.
5. Nerenz RD, Bell JR, de Oca NM, et al. Analytical and clinical evaluation of the NOWDiagnostics ADEXUSdx human chorionic gonadotropin test using whole blood. *J App Lab Med.* 2016;1:67-76.

In reply to: “Inappropriate Off-label Use of a Qualitative, Point-of-care hCG Device”

Michael Gottlieb, MD
Kristopher Wnek, MD
Jordan Moskoff, MD
Errick Christian, MD
John Bailitz, MD

John H. Stroger Hospital of Cook County, Department of Emergency Medicine,
Chicago, Illinois

Section Editor: Mark I. Langdorf, MD, MHPE

Submission history: Submitted September 28, 2016; Accepted October 11, 2016

Electronically published December 6, 2016

Full text available through open access at http://escholarship.org/uc/uciem_westjem

DOI: 10.5811/westjem.2016.10.32675

[West J Emerg Med. 2017;18(2)326.]

Dear Dr. Robert D. Nerenz, Dr. Ann M. Gronowski, and Dr David G. Grenache,

Thank you for your comments regarding our recently published article describing a reduction in turnaround time achieved by the substitution of whole blood for urine on a qualitative point-of-care hCG device.¹ We appreciate the insights and comments noted by Robert D. Nerenz, Ann M. Gronowski and David G. Grenache. The authors of this letter highlight the importance of multiple validation studies prior to routine implementation of non-FDA approved devices. We also agree with this and would like to highlight that the primary purpose of our study was to determine whether the substitution of whole blood for urine would decrease turnaround time, with the potential to reduce risks associated with delayed diagnoses of ectopic pregnancies, as well as expediting necessary imaging and treatment options that would be contingent upon pregnancy status.

While our study does support prior literature demonstrating similar accuracy between whole blood and urine for point-of-care hCG testing,² our study clearly emphasizes that further study is necessary prior to routine acceptance. One of the primary goals of our article was to justify and encourage further study into this application in order to appropriately validate it for routine clinical use.

At the time of our study, there were no FDA-approved point-of-care hCG devices that could utilize whole blood. We were excited to hear of the FDA approval of two alternate point-of-care hCG devices for use with whole blood. While our study was the first to provide evidence of an advantage in turnaround times

when using whole blood in place of urine, we look forward to further studies to determine whether similar results will be seen with these newer devices.

Address for Correspondence: Michael Gottlieb, MD, John H. Stroger Hospital Cook County, Department of Emergency Medicine, 1900 W Polk St. 10th Floor, Chicago, IL 60612. Email: michaelgottliebmd@gmail.com.

Conflicts of Interest: By the WestJEM article submission agreement, all authors are required to disclose all affiliations, funding sources and financial or management relationships that could be perceived as potential sources of bias. The authors disclosed none.

Copyright: © 2017 Gottlieb et al. This is an open access article distributed in accordance with the terms of the Creative Commons Attribution (CC BY 4.0) License. See: <http://creativecommons.org/licenses/by/4.0/>

REFERENCES

1. Gottlieb M, Wnek K, Moskoff J, et al. Comparison of result times between urine and whole blood point-of-care pregnancy testing. *West J Emerg Med.* 2016;17:449-53.
2. Fromm C, Likourezos A, Haines L, et al. Substituting whole blood for urine in a bedside pregnancy test. *J Emerg Med.* 2012;43(3):478-82.