

Images in Emergency Medicine: Pacemaker Extrusion Causing Chest Pain

Sanjay Arora, MD

LAC+USC Medical Center, Department of Emergency Medicine

Supervising Section Editor: Mark I. Langdorf MD, MHPE

Submission history: Submitted May 23, 2007; Revision Received June 19, 2007; Accepted August 3, 2007.

Reprints available through open access at www.westjem.org

[WestJEM. 2008;9:126.]



Figure.

A 55-year-old Hispanic male had a pacemaker placed in Mexico approximately one year prior to presenting to the Emergency Department. He noticed minor discomfort in his left chest one month earlier but did not see a physician. The discomfort steadily increased and he saw a small piece of metal poking through the skin. He assumed it was a staple or something minor related to the surgery; however, it gradually increased in size over the next few weeks until he realized it was the pacemaker itself eroding through his chest wall. A cardiology consultation was called, and an EKG showed

that the pacemaker was still functioning normally. He was admitted with a plan for operative repair.

Pacemaker erosion or extrusion has been reported in 0.9% of patients receiving the device.¹ The two main causes are infection and pressure necrosis.^{1,2,3} Infection has been shown to be reduced by antibiotic treatment during the peri-placement period, and pressure necrosis appears to be influenced largely by the size of the device, complexity of the connections and technical skill with which the pocket is created.^{1,2} After extrusion, the pacemaker should be considered contaminated and removed.

Address for correspondence: Sanjay Arora, MD Department of Emergency Medicine, LAC + USC Medical Center, 1200 N. State Street, Room 1011, Los Angeles, CA 90033. Email: sanjayar@usc.edu

REFERENCES

1. Shapiro M, Hanon S, Schweitzer P. A Rare, Late Complication after Automated Implantable Cardioverter-Defibrillator Placement. *Indian Pacing and Electrophysiology Journal*. 2004; 213-216.
2. Oneglia C, Gardini A, Benedini G, Rusconi C. Extrusion of a Regularly Stimulating Pacemaker. *International Journal of Cardiology*. 2007; 114:382-383.
3. Parsonnet V, Trivedi A. Pacemaker Extrusion. *Journal of the American Heart Association*. 2000; 102:1192.